

SMALL ANTENNAS—AN OVERVIEW

This article discusses state-of-the art technology and ongoing research and new developments in small antennas and data transmission links which have been designed to provide EMI measurements that would not be possible, or would be severely restricted with the use of standard antennas. It is intended that the treatment provided herein be somewhat superficial, so that an in-depth knowledge of antenna theory is not necessary to digest its content. Accordingly, the article is addressed to the professional who must deal with all engineering phases of a product, and must make sound engineering judgments based on enough information to be cognizant of the engineering problems associated with these phases.

Electromagnetic field measurements have traditionally been made with standard relatively large antennas that have been fabricated without critical attention to minimization of the sensing element or means of transmitting the received data so as not to perturb the measured field. This is not to imply that these factors were considered trivial in designing antennas. Rather, the intended use of the antennas did not justify the expense of attempting to miniaturize and provide non-perturbing data links, and in any case, the technology did not exist which could provide the means to accomplish these desirable designs.

In recent years, new technology has emerged which has placed a higher demand on the antenna design engineer to miniaturize. The factors causing this have been higher frequencies to be measured, smaller electronic units with smaller voltages and currents, and recent trends by federal agencies to consider susceptibility of individual equipment, rather than only addressing EMI emissions from intentional or inadvertent transmitting units (74).

The higher frequencies require smaller antennas due to their shorter wavelengths, the smaller electronic units require accessibility of antennas, smaller voltages and currents require close-in measurements near critical components, and the susceptibility trend will require that hardening techniques be designed into equipment that may receive undesired signals. These hardening techniques require that the electromagnetic fields be known within the equipment, whether these fields are internally or externally generated.

The types of equipment that are in need of design engineering that incorporate EMI analysis and prototype measurements are innumerable. They can range from the automobile electronic ignition and radar braking systems, to home computer games and television sets, to elaborate industrial computers and medical diagnostic equipment. All share the common potential of susceptibility to EMI, which if not thoroughly considered, can result in significant changes to production equipment. These changes may only become evident after the equipment is placed into use by the customer, and does not perform as expected due to EMI problems.

Antennas to be Considered

There are many definitions of antennas in terms of "long," "short," "small," "fat," "thin," "linear," and infinitum, with mathematical expressions for determining the fields of these antennas which may or may not be calculable. The fields are also dependent on near zone and far

zone conditions and relative sizes of transmitting and receiving antennas, ground planes, and reflecting or absorbing materials. Even an elementary treatment of the various complications in antenna parameters is far beyond the scope of this article, and is deferred to the numerous published references on antenna theory. This overview therefore consists of a cursory examination of "small" antennas that have been proven in tests, and are much less than a wavelength in their largest dimension measured from the input terminals. It will also consist of a look at ongoing and future research into antennas that provide a reasonable degree of confidence in providing accurate field measurements. Another antenna, the CAVITENNA, is discussed in the Susceptibility Sources section of this issue of ITEM.

Electromagnetic Fields of Interest

There are many different types of electromagnetic field measurements, depending on the type of sensor used. These include E-Field, H-Field, power density, current density, energy, potential, flux density, time derivative functions of these units, etc. For most EMC applications, the field parameters of interest are the E-Field (V/m) and the H-Field (A/m). Both must be measured to define the electromagnetic field in near field conditions. When far-zone conditions apply, either can be calculated from measuring the other by using the free space intrinsic impedance relationship $E = ZH$. The near zone presents complex conditions and both the E-Field and H-field must be measured to quantify the electromagnetic field. Knowledge of the E-Field and H-Field will allow the conversion into other units if the permittivity, conductivity, and permeability of the medium are known, and are uniform. Sensors have been fabricated to display these units and others directly; however, most are used in special applications, such as nuclear Electromagnetic Pulse (EMP), and geophysical measurements. For these reasons, the antennas considered in this paper will be restricted to E-Field and H-Field antennas, or sensors, which can be considered special forms of antennas.

Basic Theory of Small Sensors

The basic geometry of the passive E-Field sensor is that of a capacitive dipole as shown in Figure 1.




Figure 1. Basic Geometry of Passive E-Field Sensor

The Norton and Thevenin equivalent circuits are illustrated in Figure 2.




Figure 2. Norton and Thevenin Equivalent Circuits for Passive E-Field Sensors

The parameters of the passive electric field sensor are:

- A_e = effective area
- l_e = effective length
- L_p = sensor inductance
- μ_o = permeability of free space

From the equivalent circuits

$$A_e = \frac{C_p l_e}{\epsilon_o} \quad \text{Equation 1}$$

and the sensor output voltage is

$$V_{out} = \frac{\epsilon_o E A_e R_L}{C_p \left(R_L + \frac{1}{j\omega C_p} \right)} \quad \text{Equation 2}$$

$$\text{if } R_L \gg \frac{1}{\omega C_p}$$

$$\text{then } V_{out} = \frac{\epsilon_o A_e E}{C_p} \quad (2)$$

Equation (2) is the desired output, and is satisfied by the appropriate choice of R_L at the frequency of interest. A shunt capacitance may be needed to "swamp" the actual probe capacitance and offset stray capacitive effects. In this case, C_p in Equation (2) is replaced by the value of the shunt capacitance, provided it is much larger than C_p .

The basic geometry of the passive H-Field sensor is that of an inductive loop shown in Figure 3.




Figure 3. Basic Geometry of Passive H-Field Sensor

The Norton and Thevenin equivalent circuits are illustrated in Figure 4.




Figure 4. Norton and Thevenin Equivalent Circuits for Passive H-Field Sensors

The parameters of the passive magnetic field sensor are:

- A_e = effective area
- l_e = effective length
- C_p = sensor capacitance
- ϵ_o = permittivity of free space

From the equivalent circuits:

$$A_e = \frac{l_e L_p}{\mu_o} \quad \text{Equation 3}$$

and the sensor output voltage is

$$V_{out} = \frac{j\omega\mu_0 A_c H R_L}{j\omega L_p + R_L} \quad \text{Equation 4}$$

$$\text{if } R_L \ll \omega L_p \quad \text{Equation 5}$$

$$V_{out} = \frac{\mu_0 A_c H R_L}{L_p}$$

As with the electric field sensor, Equation (5) is the desired output for the H-Field sensor, and appropriate loading must be used to obtain this output. These relationships govern the response of small passive receiving antennas. There are, perhaps, as many shapes that can be constructed using these basic equations as there are "standard" antenna shapes and configurations. However, as the shapes become more complicated, the mathematical treatments of their performance can become extremely difficult, if not impossible, to handle. A number of sensors with relatively simple geometries have been fabricated and tested with good results. Sketches of some of these sensors are shown in Figure 5 with a typical functional block diagram of the sensor unit, data link and receiver.

Sensors 1 through 5 of Figure 5 are E-Field "surface" sensors. These measure the component of the E-Field normal to the surface over which the sensor is placed. Sensor 6 is an H-Field "surface" sensor. It measures the tangential magnetic field at the surface over which the sensor is placed. Sensors 7 through 11 are E-Field "free field" sensors, and sensor 12 is an H-Field "free field" sensor. A brief explanation of each sensor is provided, with the numbers corresponding to those in Figure 5.

1. Hemispherical Monopole over a ground plane. Its parameters are essentially the same as that of the spherical dipole with the exception that one of the hemispheres is replaced by the finite ground plane to obtain field measurements at a surface.
2. Parallel Plate (MF - VHF, approximately 1 MHz - 200 MHz) surface field sensor measures the normal component of the E-Field to the surface over which it is positioned. It is used in the lower frequency application where a cylindrical monopole would be too long to be of practical use.
3. Cylindrical Monopole (MF - UHF, approximately 10b MHz - 1 GHz) surface field sensor measures the normal component of the E-Field to the surface over which the sensor is positioned. The data transmission is coaxial cable, should be under the ground plane, or as close to it as possible, to avoid perturbing the measured field.
4. Triangular Monopole (UHF - SHF, approximately 1 GHz - 18 GHz) surface field sensor is a higher frequency version of the cylindrical monopole.
5. Conical Monopole (same as number 4.).
6. Shielded Half-Loop (VHF - UHF, approximately 200 MHz - 1 GHz) surface field sensor, measures magnetic fields tangential to the surface over which the sensor is placed. The gap reduces undesired E-Field response.
7. Spherical Dipole (ELF - MF, DC to approximately 1 MHz) free field sensor allows placement of an electronic circuit within its volume which processes the measured data into a frequency modulated signal, which is transmitted via fiber optic link. The spherical shape allows relatively easy analysis of the sensor and the fiber optic cable isolates the sensor.
8. Single Dipole (MF - UHF, approximately 100 kHz - 2 GHz) free field sensor measures the free field at relatively high frequencies. Isolation and data transmission are accomplished via high resistance transmission lines.
9. Orthogonal Dipole parameters are essentially the same as for the single dipole, with the exception that three dipoles are mounted orthogonally to provide isotropicity.
10. Single Bow-Tie (UHF - SHF, approximately 1 GHz - 18 GHz) free field sensor is a higher frequency version of the single dipole.
11. Conical Dipole (same as number 10).
12. Single Loop (MF - VHF, approximately 1 MHz - 200 MHz) free field sensor measures the magnetic free field. Isolation and data transmission are accomplished via high resistance transmission lines.




Figure 5. Existing Sensors — Block Diagram of Sensor Unit, Data Link and Receiver

Inherent Problems of Small Sensors

The sensors of today's technology are not immune to problems. Some of these problems are:

1. **Transmission data.** There are three basic means to transmit the data from the sensor to the receiver; high resistance carbon impregnated plastic lines, coaxial cable, and fiber optic links. Limitations are described below.
 - a) High resistance transmission lines—work only above 100 kHz, are subject to changes with vibration, and require that the measured signal be rectified in the sensor.
 - b) Coaxial cables—must be oriented parallel to the measured field or mounted on a ground plane.
 - c) Fiber optics—require an optical transmitter at the sensor location.
2. **Fabrication difficulty**—Not all probes of this type lend themselves to ease of fabrication—small changes in tolerance can lead to erroneous readings.
3. **Isotropy**—difficult to achieve, and must be accomplished by physically changing the orientation of the sensor, with the exception of sensor 9.
4. **Sensitivity**—decreases with size.
5. **Unwanted E-Field response** in magnetic sensors and unwanted H-Field response in electric sensors.

Although these problems are inherent in passive sensors of this type, attention to good engineering design and practice has alleviated many of these problems in sensors currently in use. Proper grounding of coaxial cables, orientation of cables, shielding, use of gaps in magnetic sensors, fiber optic data links, careful machining of parts, and the use of highly stable components have all been employed to provide a high degree of accuracy in field measurements. This accuracy, however, does approach a limit, and researchers are constantly striving for improvement in these types of sensors and the development of new sensor concepts.

New Sensor Concepts

Researchers have continued to explore new concepts in sensor technology. Examples of these concepts are glass substrate, cryogenic, Hall effect, resonant scatterers, and others. These sensors are in various stages of development in government and private research programs and are not further discussed in this paper, since most have rather specialized applications.

Although these sensors may have their rightful place in EMC, one concept has recently emerged that appears to provide a completely new method of EM measurement that may have almost unlimited application. This concept employs a jacketed fiber optic cable as a sensing element

with extremely high sensitivity for electric and magnetic fields. The fiber sensor works by magnetostrictive or electrostrictive jacketing of the light-carrying cable. When the cable is appropriately jacketed, the electric or magnetic field causes an elongation of the cable, which changes the phase of the light transmitted through the fiber. This phase change is then converted by the use of an interferometer into the electric or magnetic field strength, as appropriate. A conceptual diagram of a fiber optic sensing system is shown in Figure 6.




Figure 6. Conceptual Diagram of Fiber Optic Sensing System

A beam of light is transmitted through a single mode fiber optic cable, which is split into a reference fiber and a jacketed fiber by bottle couplers. The resultant two light beams are identical in phase. When the jacketed fiber is stressed by a magnetic or electric field, an elongation of the optical fiber results. This causes a phase change in this fiber. The phase of the reference signal is compared with that of the signal in the jacketed cable in an interferometer and the phase differential is converted into the appropriate field strength units.

The upper theoretical frequency limit of the fiber optic sensor is thought to be about 500 kHz. This limit is primarily based on the physical ability of the jacketing material to respond at a fast rate. Considering the enormous strides made in other fiber optic technology, it seems reasonable to believe that ongoing research will provide non-perturbing fiber optic sensors for the entire frequency spectrum. The sensitivity of the fiber optic sensor already betters that of conventional passive sensors by several orders of magnitude.

BIBLIOGRAPHY

1. "A Broadband Miniature Isotropic Electric Field Measurement System," H. Bassen, *1975 IEEE International EMC Symposium Record*, IEEE/75/CH1002-5 EMC, San Antonio, Texas, pp. 1-5 (5B11a), 1975.
2. "A Comparison of Electrically Short Bare and Insulated Probes for Measuring the Local Radio Frequency Electric Field in Biological Systems," G.H. Smith, *IEEE Trans. on Biomedical Engineering*, Vol. BME-22, No. 6, November 1975.
3. "A Critical Study of the Characteristics of Broadcast Antennas as Affected by Antenna Current Distribution," G.H. Brown, *Proc. I.R.E.* 24, 48, 1936.
4. "A Miniature, Optically Isolated Hemispherical Probe Over a Finite Ground Plane for Precision Surface Electric Field Measurements," M.J. Johnson, *Proc. 1981 IEEE International Symposium on Electromagnetic Compatibility*, 81CH1675-8, pp. 341-344, August 18-20, 1981.
5. "A New Approach to Electromagnetic Field Strength Measurements in Shielded Enclosures," H.A. Mendez, *Technical Report TR 02.430*, IBM Corp., June, 1968.
6. "A New Near-Zone Electric Field Strength Meter," F.M. Greene, *NBS Technical Note 345*, National Bureau of Standards, November, 1966.
7. "A Non-obtrusive Remote Indicating Instrument for Probing EM Fields Having Potentially Hazardous Levels," D.R. Kerns and E.L. Bronaugh, *Proc. IEEE International Symposium on Electromagnetic Compatibility*, Atlanta, Ga., pp. 133-136, June 20-22, 1978.
8. "A Note On The Mutual Impedance of Antennas," C.W. Harrison, Jr., *J. Appl. Phys.* 14, 306, 1943.
9. "A Small, Accurate, Optically Isolated Electric Field Probe," R.J. Spiegel, D.R. Kerns, E.H. Cooper and E.L. Bronaugh, *Proc. 1979 IEEE Power Systems and Apparatus (PES) Summer Meeting*, Paper A 79 507-5, Vancouver, B.C., July 1979.
10. "A Small ELF Electric Field Probe," A.R. Valentino, *Proc. IEEE International Symposium on Electromagnetic Compatibility*, pp. 265-270, July 1972.
11. "A Theory of Receiving Aerials Applied to the Re-Radiation of an Electromagnetic Horn," D. Midgely, *Proc. IEEE*, 108B, pp. 654-649, 1961.
12. *Advanced Antenna Theory*, S.A. Schelkunoff, John Wiley & Sons, Inc., New York, 1952.
13. "An Alternative Method of Solving Hallen's Integral Equation," R. King, *J. Appl. Phys.* 24, 140-147, 1953.
14. "An Approximate Representation of the Electromagnetic Field in the Vicinity of a Symmetrical Radiator," C.W. Harrison, Jr., *J. Appl. Phys.* 15, 544, 1944.
15. "An Equivalent-Charge Method for Defining Geometries of Dipole Antennas," C.E. Baum, Air Force Weapons Lab, *Sensor and Simulation Note 72*, Jan. 1969.
16. "An Improved Optically Isolated ELF Electric Field Sensor," E.L. Bronaugh, R.J. Spiegel, E.H. Cooper and D.R. Kerns, *1979 IEEE International Electromagnetic Compatibility Symposium Record*, IEEE 79 CH1383-9 EMC, pp. 429-434, October, 1979.
17. "An Isolated Antenna System for Shielded Enclosure Measurements, 20-200 Mhz," G.N. Van Steenberg, E.L. Bronaugh and W.C. Dolle, Southwest Research Institute, 1971 EMC International Symposium, Philadelphia, Pa., July 1971.
18. "Analysis of the High Impedance Field in Susceptibility Tests," W.A. Stirrat, U.S. Army ECOM, *Proc. 9th Tri-Service Conf. on Electromagnetic Compatibility*, pp. 747-774, 1963.
19. *Antenna Analysis*, E.A. Wolff, John Wiley & Sons, Inc., New York, 1966.
20. "Antenna Model for RFI Work," William A. Stirrat, U.S. Army ECOM, *Proc. 10th Tri-Service Conf. on Electromagnetic Compatibility*, pp. 578-591, 1964.
21. *Antenna Theory and Practice*, S.A. Schelkunoff and H.T. Friis, John Wiley & Sons, Inc., New York, 1952.
22. *Antennas*, J.D. Kraus, McGraw-Hill Book Co., New York, 1950.
23. *Antennas: Theory and Design*, S.A. Schelkunoff and H.T. Friis, John Wiley & Sons, Inc., New York, 1952.
24. "Calibration Principles and Procedures for Field Strength Meters (30 Hz to 1 GHz)," H.E. Taggart, et al., *NBS Technical Note 370*, National Bureau of Standards, Radio Standards Engineering Division, Boulder, Co., March 1969.
25. "Capacitance and Equivalent Area of a Spherical Dipole Sensor," R.W. Latham and K.S.H. Lee, Air Force Weapons Laboratory, *EM Pulse, Sensor, and Simulation Notes*, Vol. 9, Note 113, June 1971.
26. "Circuit Relations in Radiating Systems and Applications to Antenna Problems," P.S. Carter, *Proc. I.R.E.* 20, 1004, 1932.
27. "Completion of the Program to Evaluate/Improve Instrumentation and Test Methods for Electroexplosive Device Safety Qualification," Final Report, P.A. Hudson, Electromagnetics Division Institute for Basic Standard, National Bureau of Standards, Boulder, Co., NBSIR-74-379, June 1974.
28. "Concerning Hallen's Integral Equation for Cylindrical Antennas," S.A. Schelkunoff, *Proc. I.R.E.* 33, 872, 1945.
29. "Correction and Supplement to Our Paper 'The Cylindrical Antenna: Current and Impedance,'" R. King and D. Middleton, *Quart. Appl. Math* 4, 199, 1946.
30. "Coupled Antennas and Transmission Lines," R. King, *Proc. I.R.E.* 31, 626, 1943.
31. *Crystal Rectifiers*, H.C. Torrey and C.A. Whitmer, Vol. 15, MIT Rad LAB Series, McGraw-Hill Book Co., p. 3, 1948.

32. "Curves of Input Impedance Change Due to Ground for Dipole Antennas," L.E. Vogler and J.L. Noble, National Bureau of Standards, Monograph 72, Boulder, Co., Jan. 31, 1964.
33. "Design of Broadband Resistive Radiation Probes," Hopfer, et al., General Microwave Corp., *IEEE Trans. Instrum. Meas.*, Vol. IM-21, No. 4, pp. 416-421, Nov. 1972.
34. "Development and Analysis of Techniques for Calibration of Kerr Cell Pulse-Voltage Measuring Systems," Cassidy, et al., NBSIR-73-403, National Bureau of Standards, Inst. for Basic Standards, Wash., D.C., COM-74-10016, Nov. 1, 1973.
35. "Development of Electric and Magnetic Near-Field Probes," F.M. Greene, *NBS Technical Note 658*, U.S. Dept. of Commerce/National Bureau of Standards.
36. "Development of Magnetic Near-Field Probes," F.M. Greene, National Bureau of Standards, Electromagnetics Division, Boulder, Co., (NIOSH) 75-127, Jan. 1975.
37. "Development of Near-Field Electric Energy Density Meter Model EDM-2," D.R. Belsher, National Bureau of Standards, Electromagnetics Division, Boulder, Co., (NIOSH) 75-140, March 1975.
38. "Development of Probes and Measurement Techniques for Automated Interference Measurements (3 to 300 MHz) in Shielded Chambers," A.H. Mills, AFAL-TR-66-323, General Dynamics Convair Division, San Diego, California, December 1966.
39. "Directional Antennas," G.H. Brown, *Proc. I.R.E.*, Vol. 25, p. 78, Jan. 1937.
40. *Earth Conduction Effects in Transmission Systems*, D.E. Sunde, Dover Publications, 1968.
41. "Electric Field Measurements Within Biological Media," A.Y. Cheung, *Proc. Symposium on Biological Effects and Measurement of Radio Frequency/Microwaves*, HEW Pub. (FDA) 77-8026, pp. 117-135, July 1977.
42. "Electromagnetic Field Probes, Final Report, July 1, 1971 - Dec. 31, 1972," W.R. Free, et al., GIT-A-135IF, DAAB07-71-C-0306, Georgia Inst. of Tech., Atlanta, Ga., AD 909090, March 1973.
43. "Electromagnetic Interference Measurement Methods—Shielded Enclosure," W.R. Free, R.D. Gibson, B.M. Jenkins, C.W. Stuckey and J.C. Toler, Final Report ECOM-02381-F, Georgia Inst. of Tech., Atlanta, Ga., Dec. 1967.
44. "Electromagnetic Interference Measurement Methodology, Communication Equipment, Final Report, Feb. 1, 1968 - May 31, 1969," W.R. Free, et al., GIT-A-1075-F, DAAB07-68-C-0189, Georgia Inst. of Tech., Atlanta, Ga., AD 696-494.
45. "Electromagnetic Radiation Meter," E.E. Aslan, *IEEE Trans. Microwave Theory Tech.*, Vol. MTT 19, No. 2, pp. 249-250, Feb. 1971.
46. "Electromagnetic Radiation Survey Meter," E.E. Aslan, Narda Microwave Corp., Plainview, N.Y., *IEEE Trans. Instrum. Meas.*, Vol. IM-19, No. 4, pp. 368-372, Nov. 1970.
47. *Electromagnetic Waves*, S.A. Schelkunoff, D. Van Nostrand Company, Inc., Princeton, N.J., 1943.
48. *Electromagnetic Waves and Radiating Systems*, E.C. Jordan, Prentice-Hall Inc., Englewood Cliffs, N.J., 1950.
49. *Electromagnetics*, J.D. Kraus, McGraw-Hill Book Co., New York, 1953.
50. "EM Probe with Fiberoptic Telemetry System," H. Bassen, et al., *Microwave Journal*, Vol. 20, No. 4, pp. 35-47, April 1977.
51. "EMC Data Collection Techniques, Test Method RE," R.M. Showers, O.M. Salati and F. Haber, *Rome Air Development Center Technical Report RADC-TR-69-406*, University of Pennsylvania, Jan. 1970.
52. "Evaluation of Electric Field Probes for Interference Measurements," K.W. Tomlinson, AFAL-TR-68-266, Air Force Avionics Lab., Wright-Patterson AFB, Ohio, Jan. 1969.
53. "Experimental and Theoretical Impedances and Admittances of Center-Driven Antennas," P.A. Kennedy and R. King, *Cruft Laboratory Technical Report No. 155*, 1953.
54. "Experimentally Determined Radiation Characteristics of Conical and Triangular Antennas," G.H. Brown and O.M. Woodward, Jr., *RCA Review*, Vol. 13, No. 4, Dec. 1952.
55. *Field Computation by Moment Methods*, R.F. Harrington, Macmillan Co., New York, 1968.
56. "Field Intensities Around a Dipole Antenna for Distances Comparable to Antenna Length," A.H. Mills and P.L. Pelland, General Dynamics Convair Division, San Diego, Calif.
57. *Fields and Waves in Modern Radio*, S. Ramo and J.R. Whinnery, 2nd ed., John Wiley & Sons, Inc., New York, 1953.
58. "Fundamental Limitations of Small Antennas," H.A. Wheeler, *Proc. I.R.E.* 35, 1479, 1947.
59. "Generation of Standard EM Fields Using TEM Transmission Cells," M.L. Crawford, National Bureau of Standards, *IEEE Transactions on Electromagnetic Compatibility*, Vol. EMC-16, No. 4, Nov. 1974.
60. "High Frequency and Microwave Field Strength Precision Measurements Seminar," National Bureau of Standards Report, Boulder, Co., March 1966.
61. "High Frequency Models in Antenna Investigations," G.H. Brown and R. King *Proc. I.R.E.* 22, 457, 1934.
62. "Improved Techniques and Instrumentation for EMC Measurements," M.L. Crawford, *Proc. IEEE International Symposium on Electromagnetic Compatibility*, Washington, D.C., pp. 369-374, July 13-15, 1976.
63. "Internal Dosimetry and External Microwave Field Measurements Using Miniature Electric Field Probes," H.I. Bassen, *Proc. Symposium on Biological Effects and Measurement of Radio Frequency/Microwaves*, HEW Pub. (FDA) 77-8026, pp. 136-151, July 1977.

64. "Liquid Crystal Mapping of Slotted Wave Guide Fields," G.E. Fanslow, Iowa St. Univ., *J. Microwave Power*, Vol. 9, No. 3, pp. 173-177, Sept. 1974.
65. "Meaningful Measurements in Shielded Enclosures," H.A. Mendez, 1969 *IEEE EMC Symposium Record*, Asbury Park, N.J., June 1969.
66. "Measurements of the Attenuation of the Electric and Magnetic Field at Points Close to the Source," W.C. Dolle and W.E. Cory, Southwest Research Institute, *IEEE Transactions on Electromagnetic Compatibility*, Vol. EMC-10, No. 3, pp. 313-319, Sept. 1968.
67. "MGL-S8(R) B-Dot Sensor Development," S.L. Olsen, *Air Force Weapons Lab Report TR-75-252*, April 1976.
68. *Microwave Transmission*, J.C. Slater, Dover Publications, 1969.
69. "Miniature ELF Electric Field Probe," M. Misakian, F.R. Kotter and R.L. Kahler, *Rev. Scientific Instrumentation*, Vol. 49, pp. 933-935, July 1978.
70. "Model Study of Electric Field Effects in Substations," S.A. Sebo, Electric Power Research Institute, EL-632, Vol. 1, Jan. 1978.
71. "Multiple Isolated Probe Electromagnetic Field Measuring System," E.L. Bronaugh, M.L. Rose, Jr. and H.W. Rethmann, *Proc. 1980 IEEE International Symposium on Electromagnetic Compatibility*, 80CH1538-8 EMC, pp. 143-150, Oct. 1980.
72. "NBS Field-Strength Standards and Measurements," F.M. Greene, *Proc. IEEE*, Vol. 55, No. 6, pp. 970-981, June 1967.
73. "No Static Field From Alternating Currents," W.A. Stirrat, U.S. Army ECOM, *IEEE Transactions on Electromagnetic Compatibility*, Vol. EMC 13, No. 1, pp. 25-26, Feb. 1971.
74. "Notices of Inquiry into RFI to Electronic Equipment," FCC General Docket 78-369.
75. "On Radiation from Antennas," S.A. Schelkunoff and C.B. Feldman, *Proc. I.R.E.* 30, 511-516, 1942.
76. "On The Theory of Biconical Antennas," C.T. Tai, *J. Appl. Phys.* 19, 1155, 1948.
77. "Principal and Complementary Waves in Antennas," S.A. Schelkunoff, *Proc. I.R.E.* 34, p. 23, 1946.
78. "Radiation Characteristics of Dipole Sources Located Inside a Rectangular Coaxial Transmission Line," J.C. Tippet, NBSIR-75-829, National Bureau of Standards, Electromagnetics Division, Boulder, Co., PB-248855, Jan. 1976.
79. "Radiation Characteristics of Electrically Small Devices in a TEM Transmission Cell," J.C. Tippet and D.C. Chang, University of Colorado, *IEEE Transactions on Electromagnetic Compatibility*, Vol. EMC-18, No. 4, Nov. 1976.
80. "Radiation Field Measurement Study," G.N. Van Steenberg, E.L. Bronaugh, O.L. Jouffray and W.C. Dolle, USAF Cryptological Depot Project 68-24, Southwest Research Institute, Oct. 1969.
81. *Radio Wave Propagation*, C.R. Burrows and S.S. Atwood, Academic Press, New York, 1949.
82. *Reference Data for Radio Engineers*, 5th ed., Howard W. Sams & Co., Inc., A Subsidiary of ITT, Indianapolis, Indiana, 1968.
83. "Relations between the Transmitting and Receiving Properties of Antennas," A.F. Stevenson, *Quart. Appl. Math* 5, 369, 1948.
84. "Sensors for Electromagnetic Pulse Measurements Both Inside and Away from Nuclear Source Regions," C.E. Baum, et al., *IEEE Trans. Electromagnetic Compatibility*, VEMC-20, No. 1, pp. 22-34, Feb. 1978.
85. "Shielded Enclosure Measurements Can Be Accurate," G.N. Van Steenberg, J.F. William and W.C. Dolle, Southwest Research Institute, 1969 *SWIEECO Record*, April 1969.
86. "Some Further Considerations for the Circular Parallel-Plate Dipole," C.E. Baum, *Air Force Weapons Lab, Sensor and Simulation Note 86*, June 1970.
87. *Static and Dynamic Electricity*, W.R. Smythe, 3rd ed., McGraw-Hill Book Co., New York, 1968.
88. "Steady-State Solutions of Electromagnetic Field Problems: I. Forced Oscillations of a Cylindrical Conductor; II. Forced Oscillations of a Conducting Sphere; III. Forced Oscillations of a Prolate Spheroid," *J. Appl. Phys.* 12, 230, 236, 241, 1945.
89. "Studies for Correlating TEMPEST Open Field Tests to Shielded Enclosure Tests," W.C. Dolle, E.L. Bronaugh, G.N. Van Steenberg and R.A. Southwick, *ESD COMSEC Report DCW-(0109)*, Southwest Research Institute, 1970.
90. "Study of Methods of Implementing Poynting Vector Measurements," W.W. Cowles, N. Farhat and K.S. Foo, *Report No. 65-16*, Moore School of Electrical Engineering, University of Pennsylvania, Dec. 31, 1964.
91. *Tables of Functions*, E. Jahnke and F. Emde, Dover Publications, 4th ed., New York, 1945.
92. *Techniques of Microwave Measurements*, C.G. Montgomery, Vol. 11 of the MIT Rad Lab Series, McGraw-Hill Book Co., 1947.

93. "The Circular Parallel Plate Dipole," C.E. Baum, Air Force Weapons Lab, *Sensor and Simulation Note 80*, June 1970.
94. "The Cylindrical Antenna; Current and Impedance," R. King, D. Middleton, *Quart. Appl. Math* 3, 302-335, 1946.
95. "The Distribution of Current Along a Symmetrical Center-Driven Antenna," R. King and C.W. Harrison, Jr., *Proc. I.R.E.* 31, 548, 1943.
96. "The Effects of Shielded Enclosure Resonances on Measurement Accuracy," W.C. Dolle, G.N. Van Steenberg and O.L. Jonffray, Southwest Research Institute, 1970 *IEEE Electromagnetic Compatibility Symposium Record*, July 1970.
97. "The Electrical Oscillations of a Perfectly Conducting Prolate Spheroid," R.M. Ryder, *J. Appl. Phys.* 13, 327, 1942.
98. "The Hot Carrier Diode, Theory, Design and Application," *Application Note 907*, Hewlett-Packard Co., California.
99. "The Impedance of Short, Long, and Capacitively Loaded Antennas With a Critical Discussion of the Antenna Problem," R. King and C.W. Harrison, Jr., *J. Appl. Phys.* 15, 170, 1944.
100. "The Loop Antenna as a Probe," H. Whiteside and R.W.P. King, *IEEE Trans. on Antennas and Propagation*, Vol. AP-12, No. 3, pp. 291-297, May 1964.
101. "The Mutual Impedance Between Adjacent Antennas," C.R. Englund and A.B. Crawford, *Proc. I.R.E.*, 1277, 1929.
102. "The Radiating Field of a Symmetrical Center-Driven Antenna of Finite Cross Section," C.W. Harrison, Jr. and R. King, *Proc. I.R.E.* 31, 693, 1943.
103. "The Self-Impedance of a Symmetrical Antenna," R. King and F.G. Blake, Jr., *Proc. I.R.E.* 30, 335, 1942.
104. *The Theory of Spherical and Ellipsoidal Harmonics*, E.W. Hobson, Cambridge University Press, Cambridge, Mass., 1931.
105. "Theory of Antennas of Arbitrary Shape and Size," S.A. Schelkunoff, *Proc. I.R.E.* 29, 493-521, 1941.
106. "Theory of Colinear Antennas," R. King, *J. Appl. Phys.* 21, 1232, 1950.
107. "Theory of Electrically Short Transmitting and Receiving Antennas," R. King, *J. Appl. Phys.* 23, 1174-1187, 1952.
108. *Theory of Linear Antennas*, R.W.P. King, Harvard University Press, Cambridge, Mass., 1956.
109. "Theory of RFI Measurements in Shielded Enclosures," J.J. Krstanski and R.D. Standley, ITT Research Institute, Chicago, Illinois.
110. "Time Domain Solutions for Electromagnetic Coupling," A. Taflove, Document No. RADC-TR-78-142, Rome Air Development Center (RBCT) Griffiss AFB, New York.
111. *Transmission Lines, Antennas, and Wave Guides*, R.W.P. King, H.R. Mimno and A.H. Wing, McGraw-Hill Book Co., New York, 1945.
112. "Transmission Theory of Spherical Waves," S.A. Schelkunoff, *Trans. A.I.E.E.* 57, 744, 1938.
113. "USAECON Contributions to Shielding Theory," William A. Stirrat, U.S. Army ECOM, *IEEE Transactions on Electromagnetic Compatibility*, Vol. EMC-10, No. 1, pp. 63-66, March 1968.
114. "Using Optics in a Broadband, Sensitive Isotropic Antenna, 14 kHz - 150 MHz," E.B. Larsen, et al., National Bureau of Standards, Electromagnetic Division, Boulder, Co., *Proc. IEEE International Symposium on Electromagnetic Compatibility*, Washington, D.C., pp. 385-389, PB-257 168, July 13-15, 1976.

This article was prepared for ITEM by M.J. Johnson, Senior Research Engineer, and E.L. Bronaugh, Manager, EMC Research, Southwest Research Institute, San Antonio, Texas.